U.S. DEPT. OF COMMERCE NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION CLIMATE MONITORING AND DIAGNOSTICS LABORATORY DIGITAL OZONESONDE CHECKLIST

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.
DATE (LOCAL): 03 0 (108 PUMP CURRENT: 102.18 30 MINUTES HI O ₃ (10) PUMP NUMBER: 22-7299 PUMP VACUUM: 23
ADD 3.0 CC CATHODE SOLUTION: WAIT 2 MINUTES: Add about 2.5 CC more Cathode Solution (2Z) (v) ADD 1.5 CC ANODE SOLUTION: RUN 20 MINUTES ON NO O ₃ Record the current after the 20 MINUTES ON NO O ₃ : = 0.245 µamps
FLIGHT PREPARATION IN LAB. DATE (LOCAL): 03/15/08 #1: 23.18 INITIALS: 51 T100 FLOWRATE TIMES: #2: 27.24 Cathode solution date written on bottle: 8/24/07 FLOWRATE #1: 29.19 sec #3: 28.23 CHANGE CATHODE SOLUTION (3cc): (v) FLOWRATE #2: 29.21 DRY AVG: 28.72 CHANGE ANODE SOLUTION (1.5cc): (Yes/No) FLOWRATE #3: 29.28 RUN ON NO O3 FOR 5 MINUTES: (v) FLOWRATE #4: 29.28 WET T100 RECORD THE NO O3 BACKGRND#1: BG1=0.004 µamps FLOWRATE #5 29.15 #1: 28.89 RUN ON 5 microamps of O3 for 10 Minutes: (v) AVERAGE T100: 29.22 #2: 28.59 RESONSE TIME SWITCH TO NO O3 AIR. RECORD: THE TIME TO DROP FROM 4 TO 1.5 µamps: 30.84 sec. RECORD: T100 FLOWRATE TIMES: *T100 Flowrate correction. 2.27% *T100 Flowrate correction. 2.27%
DAY OF FLIGHT @ THE LAUNCH SITE. FLIGHT NUMBER: HU499 GMT DATE : 03/15/08 LOCAL DATE: 03/15/06 GMT LAUNCH TIME: 19:15 LOCAL TIME: 13:15
BALLOON TYPE 1200 Gram: Kaymont Scientific Sales (√one) O ₃ BACKGROUND (μamps from F9 key): 0,004
VAISALA NUMBER (9 digit): \\8921054\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\
REMARKS:

*T100 flow corr (%) = $[(WET/DRY)-1.0] \times 100$

weighoff = ____ grams