

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # AW512

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 5/31/08 PUMP CURRENT: 102 30 MINUTES HI O₃ (v)
INITIALS: SK PUMP PRESSURE: >11 5 MINUTE NO O₃ (v)
PUMP NUMBER: 227526 PUMP VACUUM: 22

ADD 3.0 CC CATHODE SOLUTION: (v) Short the cell leads: (v)
WAIT 2 MINUTES: (v) Add about 2.5 CC more Cathode Solution (2Z) (v)
ADD 1.5 CC ANODE SOLUTION: (v) Place Instrument inside plastic bag: (v)
RUN 20 MINUTES ON NO O₃ (v) Store inside Styrofoam flight box: (v)
Record the current after the 20 MINUTES ON NO O₃: = 0.263 μamps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 6/7/08
INITIALS: SK
Cathode solution date written on bottle: 8/24/07
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: **BG1**= 0.029 μamps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:
FLOWRATE #1: 28.66 sec
FLOWRATE #2: 28.67
FLOWRATE #3: 28.73
FLOWRATE #4: 28.55
FLOWRATE #5: 28.66
AVERAGE T100: 28.65

DRY T100

#1: _____
#2: _____
#3: _____
DRY AVG: _____

WET T100

#1: _____
#2: _____
#3: _____
WET AVG: _____

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μamps: 32.14 sec.

RECORD: ROOM TEMP (C) 25 ROOM REL. HUMID. (%) 60

RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 2 %

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: AW512

GMT DATE: 6/7/08

LOCAL DATE: 6/7/08

GMT LAUNCH TIME: 17:48:45

LOCAL TIME: 12:48:45

BALLOON TYPE _____ Gram : Kaymont Scientific Sales _____ (v one)

O₃ BACKGROUND (μamps from F9 key): _____

VAISALA NUMBER (9 digit): 219110349

SKY CONDITIONS: clear

SURFACE PRESSURE: _____

SURFACE TEMP. (C): _____

SURFACE HUMIDITY : _____

~ BURST PRESSURE (mb) : 30.41 km

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100