

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # HV541

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 11/11/08
INITIALS: B
PUMP NUMBER: 227696

PUMP CURRENT: 94.83
PUMP PRESSURE: 211
PUMP VACUUM: >22

30 MINUTES HI O₃ (v)
5 MINUTE NO O₃ (v)

ADD 3.0 CC CATHODE SOLUTION: (v)
WAIT 2 MINUTES: (v)
ADD 1.5 CC ANODE SOLUTION: (v)
RUN 20 MINUTES ON NO O₃: (v)
Record the current after the 20 MINUTES ON NO O₃: = 593 μ amps

Short the cell leads: (v)
Add about 2.5 CC more Cathode Solution (2Z): (v)
Place Instrument inside plastic bag: (v)
Store inside Styrofoam flight box: (v)

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 11/15/08
INITIALS: YR & SL

Cathode solution date written on bottle: 7/16/08
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)

RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: BG1=0.067 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:

FLOWRATE #1: 29.10 sec
FLOWRATE #2: 28.99
FLOWRATE #3: 28.95
FLOWRATE #4: 28.93
FLOWRATE #5: 29.00

AVERAGE T100: 29.994

DRY T100

#1: 28.81
#2: 28.73
#3: 28.60

DRY AVG: 28.71

WET T100

#1: 28.99
#2: 28.99
#3: 28.98

WET AVG: 28.98

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 24.60 sec.

RECORD: ROOM TEMP (C) 18°C ROOM REL. HUMID. (%) 34%

RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction: 94%

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: HV541
GMT DATE: 11/15/08
GMT LAUNCH TIME: 19:23

LOCAL DATE: 11/15/08
LOCAL TIME: 13:23

BALLOON TYPE 1200 Gram : Kaymont Scientific Sales (v one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 188322952
SURFACE PRESSURE: _____
SURFACE TEMP. (C): _____
SURFACE HUMIDITY: _____

SKY CONDITIONS: Partly cloudy & windy (ca 12 mph) from NW
~ BURST PRESSURE (mb): 4.9970
35.3 km

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100