

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # HU569

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 5/9/09
INITIALS: WC
PUMP NUMBER: 228121

PUMP CURRENT: 85.13
PUMP PRESSURE: 11
PUMP VACUUM: 24

30 MINUTES HI O₃ (v)
5 MINUTE NO O₃ (v)

ADD 3.0 CC CATHODE SOLUTION: (v) Short the cell leads: (v)
WAIT 2 MINUTES: (v) Add about 2.5 CC more Cathode Solution (2Z) (v)
ADD 1.5 CC ANODE SOLUTION: (v) Place Instrument inside plastic bag: (v)
RUN 20 MINUTES ON NO O₃: (v) Store inside Styrofoam flight box: (v)
Record the current after the 20 MINUTES ON NO O₃: = 0.414 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 5/16/09
INITIALS: WC
Cathode solution date written on bottle: 4/17/09
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: **BG1**= 0.026 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:
FLOWRATE #1: 28.50 sec
FLOWRATE #2: 28.39
FLOWRATE #3: 28.39
FLOWRATE #4: 28.41
FLOWRATE #5: 28.49
AVERAGE T100: 28.43

DRY T100
#1: 28.19
#2: 28.04
#3: 28.17
DRY AVG: 28.13
WET T100
#1: 28.72
#2: 28.68
#3: 28.65
WET AVG: 28.68

RESPONSE TIME

SWITCH TO NO O₃ AIR.
RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 23.59 sec.
RECORD: ROOM TEMP (C) 24 ROOM REL. HUMID. (%) 67
RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 1.95%

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: HU569
GMT DATE: 5/16/09 LOCAL DATE: 5/16/09
GMT LAUNCH TIME: 17:56:20 LOCAL TIME: 12:56:20

BALLOON TYPE 1200 Gram: Kaymont Scientific Sales (v one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 517104214
SURFACE PRESSURE: _____
SURFACE TEMP. (C): _____
SURFACE HUMIDITY: _____

SKY CONDITIONS: Partly Cloudy
~ BURST PRESSURE (mb): 9.793
Altitude: 31.37 km

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100