

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # HU579

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 7/11/09
INITIALS: B
PUMP NUMBER: 228230

PUMP CURRENT: 77.40
PUMP PRESSURE: >10
PUMP VACUUM: 22

30 MINUTES HI O₃ (v)
5 MINUTE NO O₃ (v)

ADD 3.0 CC CATHODE SOLUTION: (v)
WAIT 2 MINUTES: (v)
ADD 1.5 CC ANODE SOLUTION: (v)
RUN 20 MINUTES ON NO O₃: (v)

Short the cell leads: (v)
Add about 2.5 CC more Cathode Solution (2Z) (v)
Place Instrument inside plastic bag: (v)
Store inside Styrofoam flight box: (v)

Record the current after the 20 MINUTES ON NO O₃: = 0.462 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 07/25/09
INITIALS: BJL
Cathode solution date written on bottle: 4/17/09
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: BG1 = 0.020 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:
FLOWRATE #1: 28.98 sec
FLOWRATE #2: 29.23
FLOWRATE #3: 29.08
FLOWRATE #4: 28.99
FLOWRATE #5: 29.23
AVERAGE T100: 29.10

DRY T100
#1: 28.04
#2: 28.26
#3: 28.04
DRY AVG: 28.44
WET T100
#1: 28.51
#2: 28.55
#3: 28.53
WET AVG: 28.53

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 27.67 sec.

RECORD: ROOM TEMP (C) 24 ROOM REL. HUMID. (%) 49

RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 1.49 %

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: 1-HU 579
GMT DATE: 07/25/2009
GMT LAUNCH TIME: 17:55

LOCAL DATE: 07/25/2009
LOCAL TIME: 12:55

BALLOON TYPE (20) Gram: Kaymont Scientific Sales (v one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 223105105
SURFACE PRESSURE: _____
SURFACE TEMP. (C): _____
SURFACE HUMIDITY: _____

SKY CONDITIONS: Windy & cloudy

~ BURST PRESSURE (mb): 32.86 km
8.13 mb

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100