

U.S. DEPT. OF COMMERCE
 NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
 CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # Hw590

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 09/26/09
 INITIALS: SL
 PUMP NUMBER: 228449

PUMP CURRENT: 90.30
 PUMP PRESSURE: 710
 PUMP VACUUM: 21

30 MINUTES HI O₃ (v)
 5 MINUTE NO O₃ (v)

ADD 3.0 CC CATHODE SOLUTION: (v)
 WAIT 2 MINUTES: (v)
 ADD 1.5 CC ANODE SOLUTION: (v)
 RUN 20 MINUTES ON NO O₃ (v)
 Record the current after the 20 MINUTES ON NO O₃: = 0.450 μ amps

Short the cell leads: (v)
 Add about 2.5 CC more Cathode Solution (2Z) (v)
 Place Instrument inside plastic bag: (v)
 Store inside Styrofoam flight box: (v)

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 10/10/09
 INITIALS: B
 Cathode solution date written on bottle: 4/17/09
 CHANGE CATHODE SOLUTION (3cc): (v)
 CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
 RUN ON NO O₃ FOR 5 MINUTES: (v)
 RECORD THE NO O₃ BACKGRND#1: BG1 = 0.040 μ amps
 RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:
 FLOWRATE #1: 28.07 sec
 FLOWRATE #2: 28.22
 FLOWRATE #3: 28.15
 FLOWRATE #4: 28.11
 FLOWRATE #5: 28.19
AVERAGE T100: 28.15

DRY T100
 #1: 27.91
 #2: 27.90
 #3: 27.86
 DRY AVG: 27.89

WET T100
 #1: 28.42
 #2: 28.47
 #3: 28.40
 WET AVG: 28.43

RESONSE TIME

SWITCH TO NO O₃ AIR.
 RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 29.47 sec.
 RECORD: ROOM TEMP (C) 22 ROOM REL. HUMID. (%) 45
 RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 1.94 %

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: Hw590
 GMT DATE: 10/10/09 LOCAL DATE: 10/10/09
 GMT LAUNCH TIME: 1756 LOCAL TIME: 1256

BALLOON TYPE 1200 Gram: Kaymont Scientific Sales (v one)

O₃ BACKGROUND (μ amps from F9 key): 0.040

VAISALA NUMBER (9 digit): 308910948
 SURFACE PRESSURE: _____
 SURFACE TEMP. (C): _____
 SURFACE HUMIDITY: _____

SKY CONDITIONS: _____

 - BURST PRESSURE (mb): _____

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100