

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT# HU 615

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 3/20/10
INITIALS: B
PUMP NUMBER: 22-8865

PUMP CURRENT: 82.05
PUMP PRESSURE: 5.1
PUMP VACUUM: Σ3

30 MINUTES HI O₃ (v)
5 MINUTE NO O₃ (v)

ADD 3.0 CC CATHODE SOLUTION: (v)
WAIT 2 MINUTES: (v)
ADD 1.5 CC ANODE SOLUTION: (v)
RUN 20 MINUTES ON NO O₃ (v)
Record the current after the 20 MINUTES ON NO O₃: ⇒ 798 μamps

Short the cell leads: (v)
Add about 2.5 CC more Cathode Solution (2Z) (v)
Place Instrument inside plastic bag: (v)
Store inside Styrofoam flight box: (v)

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 4/3/2010
INITIALS: SL

Cathode solution date written on bottle: 10/14/09
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: BG1=0.028 μamps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:

FLOWRATE #1: 28.16 sec
FLOWRATE #2: 28.23
FLOWRATE #3: 28.23
FLOWRATE #4: 28.23
FLOWRATE #5: 28.28

AVERAGE T100: 28.226

DRY T100

#1: 27.59
#2: 27.65
#3: 27.61
DRY AVG: 27.62

WET T100

#1: 28.26
#2: 28.33
#3: 28.32
WET AVG: 28.31

RESPONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μamps: 28.85 sec.

*T100 Flowrate correction 2.46%

RECORD: ROOM TEMP (C) 22 ROOM REL. HUMID. (%) 34

RECORD: 5 - T100 FLOWRATE TIMES:

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: HU 615
GMT DATE: 4/3/10
GMT LAUNCH TIME: 18:06:26

LOCAL DATE: 4/3/10
LOCAL TIME: 13:06:26

BALLOON TYPE 1200 Gram: Kaymont Scientific Sales (none)

O₃ BACKGROUND (μamps from F9 key): 0.028

VAISALA NUMBER (9 digit): 148752840

SKY CONDITIONS: overcast

SURFACE PRESSURE: _____

SURFACE TEMP. (C): _____

SURFACE HUMIDITY: _____

* BURST PRESSURE (mb): 8.962 at 31.48/4

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100