

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # H4621

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 5/1/2010
INITIALS: WTC
PUMP NUMBER: 228880

PUMP CURRENT: 103.62
PUMP PRESSURE: 511
PUMP VACUUM: 21

30 MINUTES HI O₃ (v)
5 MINUTE NO O₃ (v)

ADD 3.0 CC CATHODE SOLUTION: (v)
WAIT 2 MINUTES: (v)
ADD 1.5 CC ANODE SOLUTION: (v)
RUN 20 MINUTES ON NO O₃: (v)

Short the cell leads: (v)
Add about 2.5 CC more Cathode Solution (2Z) (v)
Place Instrument inside plastic bag: (v)
Store inside Styrofoam flight box: (v)

Record the current after the 20 MINUTES ON NO O₃: = 0.480 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 5/5/2010
INITIALS: WTC

Cathode solution date written on bottle: 10/14/2009
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (v) (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: BG1 = 0.030 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:

FLOWRATE #1: 28.37 sec
FLOWRATE #2: 28.11
FLOWRATE #3: 28.29
FLOWRATE #4: 28.28
FLOWRATE #5: 28.29

AVERAGE T100: 28.26

DRY T100

#1: 27.89
#2: 27.83
#3: 27.85
DRY AVG: 27.85

WET T100

#1: 28.33
#2: 28.49
#3: 28.34
WET AVG: 28.38

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 52.67 sec.

RECORD: ROOM TEMP (C) 23 ROOM REL. HUMID. (%) 33

RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction 1.90 %

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: H4621
GMT DATE: 5/5/2010
GMT LAUNCH TIME: 17:34:54

LOCAL DATE: 5/5/2010
LOCAL TIME: 12:34:54

BALLOON TYPE 1200 Gram: Kaymont Scientific Sales (none)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 148752755
SURFACE PRESSURE: _____
SURFACE TEMP. (C): _____
SURFACE HUMIDITY: _____

SKY CONDITIONS: Clear

~ BURST PRESSURE (mb): _____

Altitude: 35.52 km

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100