

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # Hu626

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 05/22/2010 PUMP CURRENT: 82.54 30 MINUTES HI O₃ (✓)
INITIALS: BSW PUMP PRESSURE: 710 5 MINUTE NO O₃ (✓)
PUMP NUMBER: 22893412D PUMP VACUUM: 21

ADD 3.0 CC CATHODE SOLUTION: (✓) Short the cell leads: (✓)
WAIT 2 MINUTES: (✓) Add about 2.5 CC more Cathode Solution (2Z) (✓)
ADD 1.5 CC ANODE SOLUTION: (✓) Place Instrument inside plastic bag: (✓)
RUN 20 MINUTES ON NO O₃ (✓) Store inside Styrofoam flight box: (✓)
Record the current after the 20 MINUTES ON NO O₃: = 0.768 μamps

FLIGHT PREPARATION IN LAB.

DRY T100

DATE (LOCAL): 6/5/2010 #1: 27.54
INITIALS: WTC #2: 27.67
Cathode solution date written on bottle: 10/14/2009 #3: 27.59
CHANGE CATHODE SOLUTION (3cc): (✓) DRY AVG: 27.41
CHANGE ANODE SOLUTION (1.5cc): (Yes/No) FLOWRATE #1: 28.69 sec
RUN ON NO O₃ FOR 5 MINUTES: (✓) FLOWRATE #2: 28.52
RECORD THE NO O₃ BACKGRND#1: BG1 = 0.026 μamps FLOWRATE #3: 28.52
RUN ON 5 microamps of O₃ for 10 Minutes: (✓) FLOWRATE #4: 28.57 WET T100
FLOWRATE #5: 28.43 #1: 28.14
AVERAGE T100: 28.58 #2: 28.09
#3: 28.29
WET AVG: 28.17

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μamps: 27.56 sec.

*T100 Flowrate correction. 2.03%

RECORD: ROOM TEMP (C) 24.4 ROOM REL. HUMID. (%) 52

RECORD: 5 - T100 FLOWRATE TIMES:

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: Hu626
GMT DATE: 6/5/2010 LOCAL DATE: 6/5/2010
GMT LAUNCH TIME: 18:01:40 LOCAL TIME: 13:01:40

BALLOON TYPE 1200 Gram: Kaymont Scientific Sales (✓ one)

O₃ BACKGROUND (μamps from F9 key): _____

VAISALA NUMBER (9 digit): 309016043
SURFACE PRESSURE: _____
SURFACE TEMP. (C): _____
SURFACE HUMIDITY: _____

SKY CONDITIONS: Partly Cloudy

~ BURST PRESSURE (mb): _____

REMARKS: Took 26 minutes to go below 0.8 μamps in prep.
Altitude: 32.09 km

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100