

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT# HU642

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 9/11/2010 PUMP CURRENT: 81.20 30 MINUTES HI O₃ (v)
INITIALS: SC BH PUMP PRESSURE: 710 5 MINUTE NO O₃ (v)
PUMP NUMBER: 229422-V2D PUMP VACUUM: 23

ADD 3.0 CC CATHODE SOLUTION: (v) Short the cell leads: (v)
WAIT 2 MINUTES: (v) Add about 2.5 CC more Cathode Solution (2Z) (v)
ADD 1.5 CC ANODE SOLUTION: (v) Place Instrument inside plastic bag: (v)
RUN 20 MINUTES ON NO O₃ (v) Store inside Styrofoam flight box: (v)
Record the current after the 20 MINUTES ON NO O₃: = 0.650 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 09/18/2010
INITIALS: BH SC
Cathode solution date written on bottle: 06/21/2010
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: BG1= 0.055 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:

FLOWRATE #1: 28.93 sec
FLOWRATE #2: 28.98
FLOWRATE #3: 29.09
FLOWRATE #4: 29.15
FLOWRATE #5: 29.19

DRY T100
#1: 27.78
#2: 27.80
#3: 27.82
DRY AVG: 27.80
WET T100
#1: 28.23
#2: 28.37
#3: 28.27
WET AVG: 28.29

AVERAGE T100: 29.07

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 32.33 sec.

RECORD: ROOM TEMP (C) 23.8 ROOM REL. HUMID. (%) 48%

RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 1.76%

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: HU642
GMT DATE: 9/18/2010 LOCAL DATE: 9/18/2010
GMT LAUNCH TIME: 18:14:46 LOCAL TIME: 13:14:46

BALLOON TYPE 1200 Gram: Kaymont Scientific Sales (v one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 128323952

SKY CONDITIONS: Clear.

SURFACE PRESSURE: _____

SURFACE TEMP. (C): _____

SURFACE HUMIDITY: _____

~ BURST PRESSURE (mb): 32.93 km

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100