

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # HL 705

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 10/8/2011 PUMP CURRENT: 89.50 30 MINUTES HI O₃ (v)
INITIALS: WTC PUMP PRESSURE: 210 5 MINUTE NO O₃ (v)
PUMP NUMBER: 229854-V20 PUMP VACUUM: 22

ADD 3.0 CC CATHODE SOLUTION: 1 (v) Short the cell leads: (v)
WAIT 2 MINUTES: (v) Add about 2.5 CC more Cathode Solution (2Z) (v)
ADD 1.5 CC ANODE SOLUTION: (v) Place Instrument inside plastic bag: (v)
RUN 20 MINUTES ON NO O₃ (v) Store inside Styrofoam flight box: (v)
Record the current after the 20 MINUTES ON NO O₃: = 0.626 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 10/8/2011
INITIALS: WTC
Cathode solution date written on bottle: 3/20/2010
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (v) (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: BG1= 0.019 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:

FLOWRATE #1: 27.92 sec
FLOWRATE #2: 28.03
FLOWRATE #3: 27.89
FLOWRATE #4: 27.97
FLOWRATE #5: 28.02

AVERAGE T100: 27.95

DRY T100

#1: 27.40
#2: 27.59
#3: 27.58
DRY AVG: 27.52

WET T100

#1: 28.07
#2: 28.30
#3: 28.23
WET AVG: 28.2

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 25.44 sec.

RECORD: ROOM TEMP (C) 24 ROOM REL. HUMID. (%) 37

RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 2.47%

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: HL705
GMT DATE: 10/8/2011 LOCAL DATE: 10/8/2011
GMT LAUNCH TIME: 17:56:56 LOCAL TIME: 12:56:56

BALLOON TYPE 1000 Gram: Kaymont (v) Scientific Sales (v) (one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 123569213

SURFACE PRESSURE: _____

SURFACE TEMP. (C): _____

SURFACE HUMIDITY: _____

SKY CONDITIONS: Partly Cloudy

~ BURST PRESSURE (mb): _____

Alt: 28.66 km

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100