

U. S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # HU709

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 10/29/2011
INITIALS: SKH
PUMP NUMBER: 227878

PUMP CURRENT: 0.00
PUMP PRESSURE: 24
PUMP VACUUM: 0.23

30 MINUTES HI O₃
5 MINUTE NO O₃

ADD 3.0 CC CATHODE SOLUTION:
WAIT 2 MINUTES:
ADD 1.5 CC ANODE SOLUTION:
RUN 20 MINUTES ON NO O₃:
Record the current after the 20 MINUTES ON NO O₃: = .771 μ amps

Short the cell leads:
Add about 2.5 CC more Cathode Solution (2Z):
Place Instrument inside plastic bag:
Store inside Styrofoam flight box:

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 11/5/2011
INITIALS: WTC
Cathode solution date written on bottle: 3/20/2010
CHANGE CATHODE SOLUTION (3cc):
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES:
RECORD THE NO O₃ BACKGRND#1: BG1 = 0.052 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes:

T100 FLOWRATE TIMES:
FLOWRATE #1: 30.08 sec
FLOWRATE #2: 30.07
FLOWRATE #3: 29.91
FLOWRATE #4: 29.93
FLOWRATE #5: 30.04
AVERAGE T100: 30.00

DRY T100
#1: 28.18
#2: 28.08
#3: 27.95
DRY AVG: 28.09
WET T100
#1: 28.27
#2: 28.54
#3: 28.43
WET AVG: 28.43

RESONSE TIME

SWITCH TO NO O₃ AIR.
RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 33.33 sec.
RECORD: ROOM TEMP (C) 16.7 ROOM REL. HUMID. (%) 41
RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 1.28 %

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: HU709
GMT DATE: 11/5/11 LOCAL DATE: 11/5/11
GMT LAUNCH TIME: 18:01:07 LOCAL TIME: 13:01:07

BALLOON TYPE 1000 Gram : Kaymont Scientific Sales (✓ one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 923313204
SURFACE PRESSURE: _____
SURFACE TEMP. (C): _____
SURFACE HUMIDITY: _____

SKY CONDITIONS: Clear
~ BURST PRESSURE (mb): _____
Alt: 30.8 km

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100