

U.S. DEPT. OF COMMERCE
 NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
 CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
 DIGITAL OZONESONDE CHECKLIST

FLT # HU7H6



INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 12/10/2011 PUMP CURRENT: 89.05 30 MINUTES HI O₃ (v)
 INITIALS: BT PUMP PRESSURE: >10 5 MINUTE NO O₃ (v)
 PUMP NUMBER: 2220640 PUMP VACUUM: 22

ADD 3.0 CC CATHODE SOLUTION: (v) Short the cell leads: (v)
 WAIT 2 MINUTES: (v) Add about 2.5 CC more Cathode Solution (2Z) (v)
 ADD 1.5 CC ANODE SOLUTION: (v) Place Instrument inside plastic bag: (v)
 RUN 20 MINUTES ON NO O₃ (v) Store inside Styrofoam flight box: (v)
 Record the current after the 20 MINUTES ON NO O₃: = 0.621 μamps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 12/24/2012 INITIALS: DKT **DRY T100**
 Cathode solution date written on bottle: 9/1/2011 T100 FLOWRATE TIMES:
 CHANGE CATHODE SOLUTION (3cc): (v) FLOWRATE #1: 30.39 sec #1: 27.51
 CHANGE ANODE SOLUTION (1.5cc): (Yes/No) FLOWRATE #2: 30.32 #2: 27.57
 RUN ON NO O₃ FOR 5 MINUTES: (v) FLOWRATE #3: 30.32 #3: 27.66
 RECORD THE NO O₃ BACKGRND#1: BG1= 0.96 μamps FLOWRATE #4: 30.61 DRY AVG: 27.58
 RUN ON 5 microamps of O₃ for 10 Minutes: (v) FLOWRATE #5: 30.47 **WET T100**
 AVERAGE T100: 30.42 #1: 28.02
 #2: 28.04
 #3: 28.08
 WET AVG: 28.05

RESPONSE TIME

SWITCH TO NO O₃ AIR.
 RECORD: THE TIME TO DROP FROM 4 TO 1.5 μamps: 30.53 sec. *T100 Flowrate correction. 1.7 %
 RECORD: ROOM TEMP (C) 22.3 ROOM REL. HUMID. (%) 19
 RECORD: 5 - T100 FLOWRATE TIMES:

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: _____
 GMT DATE : _____ LOCAL DATE: _____
 GMT LAUNCH TIME : _____ LOCAL TIME: _____

BALLOON TYPE _____ Gram : Kaymont _____ Scientific Sales _____ (√ one)

O₃ BACKGROUND (μamps from F9 key): _____

VAISALA NUMBER (9 digit): 165711534 SKY CONDITIONS: _____
 SURFACE PRESSURE: _____
 SURFACE TEMP. (C): _____
 SURFACE HUMIDITY : _____ ~ BURST PRESSURE (mb) : _____

REMARKS: _____

