

U.S. DEPT. OF COMMERCE
 NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
 CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
 DIGITAL OZONESONDE CHECKLIST

FLT # H4717

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 2/17/2011 PUMP CURRENT: 85.64 30 MINUTES HI O₃ (v)
 INITIALS: AKA PUMP PRESSURE: 211 5 MINUTE NO O₃ (v)
 PUMP NUMBER: 2220641 PUMP VACUUM: 22

ADD 3.0 CC CATHODE SOLUTION: (v) Short the cell leads: (v)
 WAIT 2 MINUTES: (v) Add about 2.5 CC more Cathode Solution (2Z) (v)
 ADD 1.5 CC ANODE SOLUTION: (v) Place Instrument inside plastic bag: (v)
 RUN 20 MINUTES ON NO O₃ (v) Store inside Styrofoam flight box: (v)
 Record the current after the 20 MINUTES ON NO O₃: = 587 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 1/17/2011
 INITIALS: AKA
 Cathode solution date written on bottle: 1/18/2011
 CHANGE CATHODE SOLUTION (3cc): (v)
 CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
 RUN ON NO O₃ FOR 5 MINUTES: (v)
 RECORD THE NO O₃ BACKGRND#1: BG1 = 0.48 μ amps
 RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:
 FLOWRATE #1: 29.26 sec
 FLOWRATE #2: 29.35
 FLOWRATE #3: 29.26
 FLOWRATE #4: 29.37
 FLOWRATE #5: 29.41
 AVERAGE T100: 29.33

DRY T100
 #1: 27.56
 #2: 27.62
 #3: 27.63
 DRY AVG: 27.60

WET T100
 #1: 28.12
 #2: 28.26
 #3: 28.09
 WET AVG: 28.16

RESONSE TIME

SWITCH TO NO O₃ AIR.
RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 30.30 sec.
RECORD: ROOM TEMP (C) 22 ROOM REL. HUMID. (%) 35
RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 2.02%

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: H4717
 GMT DATE: _____ LOCAL DATE: 12/17
 GMT LAUNCH TIME: 1:00 LOCAL TIME: 1:00

BALLOON TYPE 1200 Gram : Kaymont Scientific Sales (v one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 8711555
 SURFACE PRESSURE: _____
 SURFACE TEMP. (C): _____
 SURFACE HUMIDITY : _____

SKY CONDITIONS: RAIN
 ~ BURST PRESSURE (mb) : 27.5 km

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100