

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # HU 718

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT. 101.23

DATE (LOCAL): 12/24/2012 PUMP CURRENT: ~~0.00~~ 30 MINUTES HI O₃: (✓)
INITIALS: AKH PUMP PRESSURE: 711 5 MINUTE NO O₃: (✓)
PUMP NUMBER: 2220635 PUMP VACUUM: ~~0.00~~ 14

ADD 3.0 CC CATHODE SOLUTION: (✓) Short the cell leads: (✓)
WAIT 2 MINUTES: (✓) Add about 2.5 CC more Cathode Solution (2Z): (✓)
ADD 1.5 CC ANODE SOLUTION: (✓) Place Instrument inside plastic bag: (✓)
RUN 20 MINUTES ON NO O₃: (✓) Store inside Styrofoam flight box: (✓)
Record the current after the 20 MINUTES ON NO O₃: = 0.507 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 01/14/2012
INITIALS: BH
Cathode solution date written on bottle: 09/08/2011
CHANGE CATHODE SOLUTION (3cc): (✓)
CHANGE ANODE SOLUTION (1.5cc): (✓) (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (✓)
RECORD THE NO O₃ BACKGRND#1: BG1=0.075 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (✓)

T100 FLOWRATE TIMES:
FLOWRATE #1: 29.37 sec
FLOWRATE #2: 28.44
FLOWRATE #3: 29.26
FLOWRATE #4: 29.29
FLOWRATE #5: 29.32
AVERAGE T100: 29.34

DRY T100
#1: 27.48
#2: 27.48
#3: 27.43
DRY AVG: 27.46
WET T100
#1: 28.03
#2: 27.93
#3: 27.98
WET AVG: 27.96

RESONSE TIME

SWITCH TO NO O₃ AIR.
RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 27.81 sec. *T100 Flowrate correction. 1.82%
RECORD: ROOM TEMP (C) 21 ROOM REL. HUMID. (%) 13
RECORD: 5 - T100 FLOWRATE TIMES:

DAY-OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: _____
GMT DATE : _____ LOCAL DATE: _____
GMT LAUNCH TIME : _____ LOCAL TIME: _____

BALLOON TYPE _____ Gram : Kaymont _____ Scientific Sales _____ (✓ one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 926307212 SKY CONDITIONS: _____
SURFACE PRESSURE: _____
SURFACE TEMP. (C): _____
SURFACE HUMIDITY : _____ ~ BURST PRESSURE (mb) : _____

REMARKS: _____