

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # HU724

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 2/21/2012 PUMP CURRENT: 28.85 30 MINUTES HI O₃ (v)
INITIALS: SMR PUMP PRESSURE: 211 5 MINUTE NO O₃ (v)
PUMP NUMBER: 2210206 PUMP VACUUM: 22

ADD 3.0 CC CATHODE SOLUTION: (v) Short the cell leads: (v)
WAIT 2 MINUTES: (v) Add about 2.5 CC more Cathode Solution (2Z) (v)
ADD 1.5 CC ANODE SOLUTION: (v) Place Instrument inside plastic bag: (v)
RUN 20 MINUTES ON NO O₃ (v) Store inside Styrofoam flight box: (v)
Record the current after the 20 MINUTES ON NO O₃: = 798 μ amps
+ 1 min.

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 02/25 DRY T100
INITIALS: BZT #1: 27.73
Cathode solution date written on bottle: 09/08/2011 T100 FLOWRATE TIMES: #2: 28.09
CHANGE CATHODE SOLUTION (3cc): (v) FLOWRATE #1: 29.79 sec #3: 27.77
CHANGE ANODE SOLUTION (1.5cc): (Yes/No) FLOWRATE #2: 29.77 DRY AVG: 27.86
RUN ON NO O₃ FOR 5 MINUTES: (v) FLOWRATE #3: 29.80
RECORD THE NO O₃ BACKGRND#1: BG1= 0.082 μ amps FLOWRATE #4: 29.77

RUN ON 5 microamps of O₃ for 10 Minutes: (v) FLOWRATE #5: 29.84 WET T100
AVERAGE T100: 29.79 #1: 28.77
#2: 28.31
#3: 28.31
WET AVG: 28.30

RESONSE TIME
SWITCH TO NO O₃ AIR.
RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 31.20 sec. *T100 Flowrate correction. 1.58%
RECORD: ROOM TEMP (C) 19.4 ROOM REL. HUMID. (%) 17
RECORD: 5 - T100 FLOWRATE TIMES:

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: _____
GMT DATE : _____ LOCAL DATE: _____
GMT LAUNCH TIME : _____ LOCAL TIME: _____

BALLOON TYPE _____ Gram : Kaymont _____ Scientific Sales _____ (v one)

O₃ BACKGROUND (μ amps from F9 key): 18

VAISALA NUMBER (9 digit): 168711643 SKY CONDITIONS: _____
SURFACE PRESSURE: _____
SURFACE TEMP. (C): _____
SURFACE HUMIDITY : _____ ~ BURST PRESSURE (mb) : _____

REMARKS: _____