

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # 44237

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 05/22/2012
INITIALS: 134
PUMP NUMBER: 7310070

PUMP CURRENT: 87.40
PUMP PRESSURE: 710
PUMP VACUUM: 24

30 MINUTES HI O₃ (v)
5 MINUTE NO O₃ (v)

ADD 3.0 CC CATHODE SOLUTION: (v)
WAIT 2 MINUTES: (v)
ADD 1.5 CC ANODE SOLUTION: (v)
RUN 20 MINUTES ON NO O₃: (v)
Record the current after the 20 MINUTES ON NO O₃: = 0.509 μ amps

Short the cell leads: (v)
Add about 2.5 CC more Cathode Solution (2Z): (v)
Place Instrument inside plastic bag: (v)
Store inside Styrofoam flight box: (v)

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 5/26/2012
INITIALS: WTC
Cathode solution date written on bottle: Sept 8 2011
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: BG1= 0.069 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:
FLOWRATE #1: 29.9 sec
FLOWRATE #2: 29.8
FLOWRATE #3: 29.8
FLOWRATE #4: 29.9
FLOWRATE #5: 29.8
AVERAGE T100: 29.85

DRY T100
#1: 29.1
#2: 29.2
#3: 29.1
DRY AVG: 29.15
WET T100
#1: 30.7
#2: 30.6
#3: 30.7
WET AVG: 30.65

RESONSE TIME

SWITCH TO NO O₃ AIR.
RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 27.2 sec.
RECORD: ROOM TEMP (C) 24 ROOM REL. HUMID. (%) 47
RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. _____ %

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: 44237
GMT DATE: 5/26/2012 LOCAL DATE: 5/24/2012
GMT LAUNCH TIME: _____ LOCAL TIME: _____

BALLOON TYPE 400 Gram : Kaymont Scientific Sales _____ (v one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 239113443 SKY CONDITIONS: clear
SURFACE PRESSURE: _____
SURFACE TEMP. (C): _____ +4.2
SURFACE HUMIDITY: _____ ~ BURST PRESSURE (mb): _____

REMARKS: _____

weighoff = _____ grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100