

U.S. DEPT. OF COMMERCE
 NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
 CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

#4739
 FLT # ~~4739~~

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 5/31/2012 PUMP CURRENT: 85.77 30 MINUTES HI O₃: (v)
 INITIALS: WTR PUMP PRESSURE: 701 5 MINUTE NO O₃: (v)
 PUMP NUMBER: 2210069 PUMP VACUUM: 22

ADD 3.0 CC CATHODE SOLUTION: (v) Short the cell leads: (v)
 WAIT 2 MINUTES: (v) Add about 2.5 CC more Cathode Solution (2Z): (v)
 ADD 1.5 CC ANODE SOLUTION: (v) Place Instrument inside plastic bag: (v)
 RUN 20 MINUTES ON NO O₃: (v) Store inside Styrofoam flight box: (v)
 Record the current after the 20 MINUTES ON NO O₃: = 0.545 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 6/9/2012
 INITIALS: WTR
 Cathode solution date written on bottle: Sept 8, 2011
 CHANGE CATHODE SOLUTION (3cc): (v)
 CHANGE ANODE SOLUTION (1.5cc): (v) (Yes/No)
 RUN ON NO O₃ FOR 5 MINUTES: (v)
 RECORD THE NO O₃ BACKGRND#1: BG1= 0.071 μ amps
 RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:
 FLOWRATE #1: 29.4 sec
 FLOWRATE #2: 29.5
 FLOWRATE #3: 29.1
 FLOWRATE #4: 29.3
 FLOWRATE #5: 29.7
 AVERAGE T100: 29.4

DRY T100
 #1: 29.5
 #2: 29.4
 #3: ~~29.9~~ 29.9
 DRY AVG: 29.6
WET T100
 #1: 30.4
 #2: 30.2
 #3: 31.0
 WET AVG: 30.5

RESONSE TIME

SWITCH TO NO O₃ AIR.
 RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 28.0 sec.
 RECORD: ROOM TEMP (C) 24 ROOM REL. HUMID. (%) 44
 RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 3.04%

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: H4739
 GMT DATE: 6/9/2012 LOCAL DATE: 6/9/2012
 GMT LAUNCH TIME: 17:46:49 LOCAL TIME: 12:46:49

BALLOON TYPE 800 Gram: Kaymont (v) Scientific Sales (v one)

O₃ BACKGROUND (μ amps from F9 key): _____

VAISALA NUMBER (9 digit): 239113444 SKY CONDITIONS: _____
 SURFACE PRESSURE: _____
 SURFACE TEMP. (C): _____ +4.5
 SURFACE HUMIDITY: _____ ~ BURST PRESSURE (mb): _____
30.7 km

REMARKS: _____