

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # HU 834

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 10/05/2013 PUMP CURRENT: 93.48 30 MINUTES HI O₃ (✓)
INITIALS: NLP PUMP PRESSURE: 210 5 MINUTE NO O₃ (✓)
PUMP NUMBER: 2224597 PUMP VACUUM: 23

ADD 3.0 CC CATHODE SOLUTION: (✓) Short the cell leads: (✓)
WAIT 2 MINUTES: (✓) Add about 2.5 CC more Cathode Solution (2Z) (✓)
ADD 1.5 CC ANODE SOLUTION: (✓) Place Instrument inside plastic bag: (✓)
RUN 20 MINUTES ON NO O₃ (✓) Store inside Styrofoam flight box: (✓)
Record the current after the 20 MINUTES ON NO O₃: = 0.716 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 10/12/2013 **DRY T100**
INITIALS: NLP #1: 29.09
Cathode solution date written on bottle: 239 #2: 29.19
CHANGE CATHODE SOLUTION (3cc): (✓) T100 FLOWRATE TIMES: #3: 29.16
CHANGE ANODE SOLUTION (1.5cc): (Yes/No) FLOWRATE #1: 29.55 sec DRY AVG: 29.15
RUN ON NO O₃ FOR 5 MINUTES: (✓) FLOWRATE #2: 29.49
RECORD THE NO O₃ BACKGRND#1: BG1 = 0.056 μ amps FLOWRATE #3: 29.66
RUN ON 5 microamps of O₃ for 10 Minutes: (✓) FLOWRATE #4: 29.57 **WET T100**
AVERAGE T100: 29.57 #1: 29.43
#2: 29.47
#3: 29.47
WET AVG: 29.41

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 25.19 sec. *T100 Flowrate correction. 1.06 %

RECORD: ROOM TEMP (C) 23.3 ROOM REL. HUMID. (%) 42

RECORD: 5 - T100 FLOWRATE TIMES:

DAY OF FLIGHT @ THE LAUNCH SITE. NLP

FLIGHT NUMBER: HU 834

GMT DATE: 10/12/2013

LOCAL DATE: 10/12/2013

GMT LAUNCH TIME: 18:00

LOCAL TIME: 13:00

BALLOON TYPE 800 Gram: Kaymont _____ Scientific Sales _____ (None) Hwoyee

O₃ BACKGROUND (μ amps from F9 key): 0.056

VAISALA NUMBER (9 digit): 21002

SURFACE PRESSURE: 992.1

SURFACE TEMP. (C): 16.9

SURFACE HUMIDITY: 75.2

SKY CONDITIONS: Mostly sunny w/a high near 82 North wind around 5 mph
- BURST PRESSURE (mb): 31.20

REMARKS:

weighoff = 1700 grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100