

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # H4838

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 10/26/2013 PUMP CURRENT: 95.29 30 MINUTES HI O₃ (✓)
INITIALS: NLP PUMP PRESSURE: >10 5 MINUTE NO O₃ (✓)
PUMP NUMBER: 2224470 PUMP VACUUM: 20

ADD 3.0 CC CATHODE SOLUTION: (✓) Short the cell leads: (✓)
WAIT 2 MINUTES: (✓) Add about 2.5 CC more Cathode Solution (2Z) (✓)
ADD 1.5 CC ANODE SOLUTION: (✓) Place Instrument inside plastic bag: (✓)
RUN 20 MINUTES ON NO O₃: (✓) Store inside Styrofoam flight box: (✓)
Record the current after the 20 MINUTES ON NO O₃: = 0.068 μ amps
0.164

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 11/9/2013
INITIALS: WTC
Cathode solution date written on bottle: Aug 13, 2013
CHANGE CATHODE SOLUTION (3cc): (✓)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (✓)
RECORD THE NO O₃ BACKGRND#1: BG1 = 0.068 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (✓)

T100 FLOWRATE TIMES:
FLOWRATE #1: 29.05 sec
FLOWRATE #2: 28.94
FLOWRATE #3: 29.14
FLOWRATE #4: 29.19
FLOWRATE #5: 29.23
AVERAGE T100: 29.11

DRY T100
#1: 28.65
#2: 28.80
#3: 28.25
DRY AVG: 28.77
WET T100
#1: 29.12
#2: 29.27
#3: 28.84
WET AVG: 29.23

RESONSE TIME

SWITCH TO NO O₃ AIR.
RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 32.05 sec.
RECORD: ROOM TEMP (C) 19 ROOM REL. HUMID. (%) 26
RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 160 %

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: H4838
GMT DATE: 11/9/2013 LOCAL DATE: 11/9/2013
GMT LAUNCH TIME: 19:04 LOCAL TIME: 13:04

BALLOON TYPE 1200 Gram: Kaymont Scientific Sales (none) Hwoyee

O₃ BACKGROUND (μ amps from F9 key): 0.068

VAISALA NUMBER (9 digit): 21114 SKY CONDITIONS: _____
SURFACE PRESSURE: 997.7 _____
SURFACE TEMP. (C): 21.9 _____
SURFACE HUMIDITY: 59.8 ~ BURST PRESSURE (mb): _____

REMARKS: _____

weighoff = 1700 grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100