

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # H4841

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 11/23/2013 PUMP CURRENT: 107.86 30 MINUTES HI O₃ (✓)
INITIALS: NLP PUMP PRESSURE: >10 5 MINUTE NO O₃ (✓)
PUMP NUMBER: 2224469 PUMP VACUUM: 19

ADD 3.0 CC CATHODE SOLUTION: (✓) Short the cell leads: (✓)
WAIT 2 MINUTES: (✓) Add about 2.5 CC more Cathode Solution (2Z) (✓)
ADD 1.5 CC ANODE SOLUTION: (✓) Place Instrument inside plastic bag: (✓)
RUN 20 MINUTES ON NO O₃: (✓) Store inside Styrofoam flight box: (✓)
Record the current after the 20 MINUTES ON NO O₃: = 1.493 μamps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 11/30/2013
INITIALS: NLP
Cathode solution date written on bottle: 239
CHANGE CATHODE SOLUTION (3cc): (✓)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (✓)
RECORD THE NO O₃ BACKGRND#1: BG1 = 0.057 μamps
RUN ON 5 microamps of O₃ for 10 Minutes: (✓)

T100 FLOWRATE TIMES:

FLOWRATE #1: 28.71 sec
FLOWRATE #2: 28.58
FLOWRATE #3: 28.54
FLOWRATE #4: 28.67
FLOWRATE #5: 28.63
AVERAGE T100: 28.63

DRY T100

#1: 29.03
#2: 29.08
#3: 29.08
DRY AVG: 29.06

WET T100

#1: 29.52
#2: 29.43
#3: 29.40
WET AVG: 29.45

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μamps: 39.63 sec.

RECORD: ROOM TEMP (C) 15.8 ROOM REL. HUMID. (%) 81

RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction: 1.34 %

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: H4841
GMT DATE: 11/30/2013 LOCAL DATE: 11/30/2013
GMT LAUNCH TIME: _____ LOCAL TIME: _____

BALLOON TYPE 1200 Gram: Kaymont _____ Scientific Sales _____ (None) Hwoyee ✓

O₃ BACKGROUND (μamps from F9 key): 0.057

VAISALA NUMBER (9 digit): 21113
SURFACE PRESSURE: 1004.2
SURFACE TEMP. (C): 13.7
SURFACE HUMIDITY: 30.5

SKY CONDITIONS: _____
~ BURST PRESSURE (mb): _____

REMARKS: _____

weighoff = 1700 grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100