

U.S. DEPT. OF COMMERCE
NATIONAL OCEANIC AND ATMOSPHERIC ADMINISTRATION
CLIMATE MONITORING AND DIAGNOSTICS LABORATORY
DIGITAL OZONESONDE CHECKLIST

FLT # HU 851

Huntsville

INITIAL PREPARATION 3-7 DAYS BEFORE FLIGHT.

DATE (LOCAL): 1/28/2014 PUMP CURRENT: 70.00 30 MINUTES HI O₃ (v)
INITIALS: WMT PUMP PRESSURE: 10.16 5 MINUTE NO O₃ (v)
PUMP NUMBER: 2723098 PUMP VACUUM: 22

ADD 3.0 CC CATHODE SOLUTION: (v) Short the cell leads: (v)
WAIT 2 MINUTES: (v) Add about 2.5 CC more Cathode Solution (2Z) (v)
ADD 1.5 CC ANODE SOLUTION: (v) Place Instrument inside plastic bag: (v)
RUN 20 MINUTES ON NO O₃ (v) Store inside Styrofoam flight box: (v)
Record the current after the 20 MINUTES ON NO O₃: = 0.145 μ amps

FLIGHT PREPARATION IN LAB.

DATE (LOCAL): 02/08/2014
INITIALS: NLP
Cathode solution date written on bottle: 239
CHANGE CATHODE SOLUTION (3cc): (v)
CHANGE ANODE SOLUTION (1.5cc): (Yes/No)
RUN ON NO O₃ FOR 5 MINUTES: (v)
RECORD THE NO O₃ BACKGRND#1: BG1 = 0.054 μ amps
RUN ON 5 microamps of O₃ for 10 Minutes: (v)

T100 FLOWRATE TIMES:

FLOWRATE #1: 28.39 sec
FLOWRATE #2: 28.47
FLOWRATE #3: 28.33
FLOWRATE #4: 28.21
FLOWRATE #5: 28.29
AVERAGE T100: 28.34

DRY T100

#1: 29.07
#2: 29.18
#3: 29.13
DRY AVG: 29.13

WET T100

#1: 29.47
#2: 29.53
#3: 29.53
WET AVG: 29.51

RESONSE TIME

SWITCH TO NO O₃ AIR.

RECORD: THE TIME TO DROP FROM 4 TO 1.5 μ amps: 27.03 sec.

RECORD: ROOM TEMP (C) 18.9 ROOM REL. HUMID. (%) 18

RECORD: 5 - T100 FLOWRATE TIMES:

*T100 Flowrate correction. 1.30%

DAY OF FLIGHT @ THE LAUNCH SITE.

FLIGHT NUMBER: HU 851
GMT DATE: 02/08/2014 LOCAL DATE: 02/08/2014
GMT LAUNCH TIME: 19:00 LOCAL TIME: 13:00

BALLOON TYPE 1200 Gram: Kaymont Scientific Sales (v one) Hwoyee

O₃ BACKGROUND (μ amps from F9 key): 0.054

VAISALA NUMBER (5 digit): 21386
SURFACE PRESSURE: 1004.5
SURFACE TEMP. (C): -3.2 C
SURFACE HUMIDITY: 74

SKY CONDITIONS: _____

~ BURST PRESSURE (mb): 33.37
ALT.

REMARKS: _____

weighoff = 1700 grams

*T100 flow corr (%) = [(WET/DRY)-1.0] X 100